


TECHNICAL DATA SHEET

Product Name: Stellant Dual Syringe Family

Catalog Numbers: **SDS-CTP-QFT**, SDS-CTP-SPK, SDS-TRF-CV, SDS-CTP-QFT-CV

Injector Series: Stellant CT Injector System



| Item | Component* | Material |
|------|------------------------------------|--|
| 1 | Plunger Assembly | Polycarbonate/Synthetic Rubber or Polycarbonate/Synthetic Rubber/Polypropylene |
| 2 | Barrel | Polyethylene Terephthalate |
| 3 | Dust Cover | Polypropylene |
| 4 | Quick Fill Tube | Low density polyethylene |
| 5 | Spike | ABS |
| 6 | Low Pressure Connector Tube | PVC/Polycarbonate/Polypropylene |
| 7 | Prime Tube | LDPE/Polyethylene |
| 8 | Transfer Set | PVC/Polycarbonate/ABS/Silicone/Polypropylene/Pallflex/Polyallomer/LDPE/Stainless Steel |
| 9 | Check Valve | Polycarbonate/Silicone |

Shelf Life: 4 years

Primary Packaging: Polystyrene tray with Tyvek® lid

Secondary Packaging: Cardboard shipper box

Sterilization Method: Ethylene Oxide (EtO), Irradiation

Latex free

MEDRAD, INC. USA
100 Global View Drive
Warrendale, PA 15086
USA
Phone: (412) 767-2400
Fax: (412) 767-4120

MEDRAD Europe B.V.
Postbus 205
6190 AE Beek
The Netherlands
Phone: +31 (0) 43-3585600
Fax: +31 (0) 43-3656598



*Note: Components shown may change according to kit configuration